

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Instrument for measuring differential pressure, flow, temperature and power through balancing valves in hydronic systems
Name and address of the applicant	IMI Hydronic Engineering AB SE-524 80 Ljung, Annelund Sweden
Name and address of the manufacturer	Same as applicant
Name and address of the factory <i>Note: When more than one factory, please report on page 2</i>	Same as applicant
Ratings and principal characteristics	TA SCOPE consist of: DpS Visio: 5.2 VDC, Class I, IP64. Internal battery: 3.2VDC Handheld unit: 5VDC, Class I, IP64. Internal battery: 3.7VDC With battery charger: 100-240V, 2A, Class II
Trademark (if any)	IMI TA
Customer's Testing Facility (CTF) Stage used	-
Model / Type Ref.	TA SCOPE
Additional information (if necessary may also be reported on page 2)	See page 2
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010+A1
As shown in the Test Report Ref. No. which forms part of this Certificate	1908887STO-001

This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB
Box 1103
SE-164 22 Kista, Sweden
Int +46 8 750 00 00

Date: 20 August 2019



Signature:


Henrik Wikström



Ref. Certif. No.

SE-86671M2

Additional information

This certificate replaces CB certificate SE-8667M1, dated 20 March 2018. A new certificate has been issued due to the merging of DpS-Visio and Handheld unit under the common name TA SCOPE and an update of the reports to edition 3.1.

The product is also in conformity with the following national differences as shown in report with ref. No. 1908887STO-001:

- CSA C22.2 No. 61010-1-12
- UL 61010-1:2012

On-site verification of the production line testing according to Annex F was excluded.

Date: 20 August 2019

Signature: